

Revised Laboratory Analytical Report

14 January 2026

Ms. Jill Douthit

Miami Public Schools
601 16th Avenue, Northwest
Miami, OK 74354



WO: E6A0125

RE: Will Rogers School Site

Enclosed are the results of analyses for samples received by the laboratory on 1/8/2026. If you have any questions concerning this report, please feel free to contact me.

Sincerely,

A handwritten signature in black ink, appearing to read "Russell Britten", is written over a light gray rectangular background.

Russell Britten
President





4619 N. Santa Fe Ave
Oklahoma City, OK 73118
405.488.2400 Phone
405.488.2404 Fax
www.etilab.com

Miami Public Schools
601 16th Avenue, Northwest
Miami, OK 74354

Project: Will Rogers School Site

Project Number: [none]

Project Manager: Ms. Jill Douthit

Reported:
01/14/26 09:19

Work Order Case Narrative

Case Narrative

The enclosed soil sample from Will Rogers School Site was received on 01/08/26 and analyzed on 01/09/26 for benzene content. There was no benzene found in the sample.

Environmental Testing, Inc.

Russell Britten, President

The results in this report apply to the samples analyzed in accordance with the chain of custody document and meet all laboratory accreditation requirements unless noted otherwise. This analytical report must be reproduced in its entirety.



E6A0125
Revised
ETI_OKC_RPT MRL_rev51.0.rpt



4619 N. Santa Fe Ave
Oklahoma City, OK 73118
405.488.2400 Phone
405.488.2404 Fax
www.etilab.com

Miami Public Schools
601 16th Avenue, Northwest
Miami, OK 74354

Project: Will Rogers School Site
Project Number: [none]
Project Manager: Ms. Jill Douthit

Reported:
01/14/26 09:19

Will Rogers School Site

E6A0125-01 (Solid) - Sampled: 01/08/26 10:30

Analyte	Result	Reporting Limit	Units	Dilution	Batch	Analyst	Analyzed	Method	Qualifiers
Volatile Organic Compounds by EPA Method 8021									
Benzene	<0.025	0.025	mg/Kg	0.987	EOA0123	WM	01/09/26 16:30	EPA 8021B 1996	
Surrogate: a,a,a-Trifluorotoluene		80 %		45.4-152	EOA0123	WM	01/09/26 16:30	EPA 8021B 1996	
Surrogate: 4-Bromofluorobenzene		101 %		41.7-151	EOA0123	WM	01/09/26 16:30	EPA 8021B 1996	

Environmental Testing, Inc.

Russell Britten, President

The results in this report apply to the samples analyzed in accordance with the chain of custody document and meet all laboratory accreditation requirements unless noted otherwise. This analytical report must be reproduced in its entirety.



E 6 A 0 1 2 5

E6A0125
Revised
ETI_OKC_RPT_MRL_rev51.0.rpt

Miami Public Schools
601 16th Avenue, Northwest
Miami, OK 74354

Project: Will Rogers School Site
Project Number: [none]
Project Manager: Ms. Jill Douthit

Reported:
01/14/26 09:19

QUALITY CONTROL

Volatile Organic Compounds by EPA Method 8021 Environmental Testing, Inc.

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Qualifiers
---------	--------	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	------------

Batch EOA0123 - EPA 5035 Soil GC

Blank (EOA0123-BLK1)

Prepared & Analyzed: 01/09/26

Benzene	<0.025	0.025	mg/Kg							
Surrogate: a,a,a-Trifluorotoluene	0.130		mg/Kg	0.1500		87	45.4-152			
Surrogate: 4-Bromofluorobenzene	0.156		mg/Kg	0.1500		104	41.7-151			

LCS (EOA0123-BS1)

Prepared & Analyzed: 01/09/26

Benzene	0.486	0.025	mg/Kg	0.4912		99	63.8-138			
Toluene	0.559	0.025	mg/Kg	0.4912		114	80.9-135			
Ethylbenzene	0.566	0.025	mg/Kg	0.4912		115	75.1-130			
Xylenes (total)	1.70	0.074	mg/Kg	1.473		116	81-125			
Surrogate: a,a,a-Trifluorotoluene	0.145		mg/Kg	0.1473		99	45.4-152			
Surrogate: 4-Bromofluorobenzene	0.174		mg/Kg	0.1473		118	41.7-151			

Matrix Spike (EOA0123-MS1)

Source: E6A0064-01

Prepared & Analyzed: 01/09/26

Benzene	0.480	0.025	mg/Kg	0.4965	0.005	96	52.3-133			
Toluene	0.539	0.025	mg/Kg	0.4965	ND	109	61.1-130			
Ethylbenzene	0.533	0.025	mg/Kg	0.4965	0.008	106	54.2-129			
Xylenes (total)	1.58	0.074	mg/Kg	1.490	0.019	105	56.8-128			
Surrogate: a,a,a-Trifluorotoluene	0.137		mg/Kg	0.1490		92	45.4-152			
Surrogate: 4-Bromofluorobenzene	0.152		mg/Kg	0.1490		102	41.7-151			

Matrix Spike Dup (EOA0123-MSD1)

Source: E6A0064-01

Prepared & Analyzed: 01/09/26

Benzene	0.485	0.025	mg/Kg	0.4995	0.005	96	52.3-133	1	20	
Toluene	0.550	0.025	mg/Kg	0.4995	ND	110	61.1-130	2	20	
Ethylbenzene	0.574	0.025	mg/Kg	0.4995	0.008	113	54.2-129	7	20	
Xylenes (total)	1.66	0.075	mg/Kg	1.499	0.019	110	56.8-128	5	20	
Surrogate: a,a,a-Trifluorotoluene	0.138		mg/Kg	0.1499		92	45.4-152			
Surrogate: 4-Bromofluorobenzene	0.153		mg/Kg	0.1499		102	41.7-151			

Environmental Testing, Inc.



Russell Britten, President

The results in this report apply to the samples analyzed in accordance with the chain of custody document and meet all laboratory accreditation requirements unless noted otherwise. This analytical report must be reproduced in its entirety.



E 6 A 0 1 2 5



4619 N. Santa Fe Ave
Oklahoma City, OK 73118
405.488.2400 Phone
405.488.2404 Fax
www.etilab.com

Miami Public Schools
601 16th Avenue, Northwest
Miami, OK 74354

Project: Will Rogers School Site
Project Number: [none]
Project Manager: Ms. Jill Douthit

Reported:
01/14/26 09:19

Certifications

Code	Description	Number	Expires
NELAP/OK	NELAP Accredited (ODEQ)	2026-090	12/31/2026
TCEQ	Texas Accredited (TCEQ)	TX-C25-00090	03/31/2026

Qualifiers and Definitions

Abbreviation	Description
DET	Analyte DETECTED
ND	Analyte NOT DETECTED at or above the reporting limit
NR	Not Reported
dry	Sample results reported on a dry weight basis
RPD	Relative Percent Difference
x	Non-Certified analyte
NA	Not Applicable
Qualifier	Description

Environmental Testing, Inc.

Russell Britten, President

The results in this report apply to the samples analyzed in accordance with the chain of custody document and meet all laboratory accreditation requirements unless noted otherwise. This analytical report must be reproduced in its entirety.



E 6 A 0 1 2 5

E6A0125
Revised
ETI_OKC_RPT_MRL_rev51.0.rpt

Page 5 of 7



Client: Miami Public Schools
Project: Will Rogers School Site

Project Manager: Ms. Jill Douthit
Project Number: [none]

Report To:
Miami Public Schools
Ms. Jill Douthit
601 16th Avenue, Northwest
Miami, OK 74354

Invoice To:
Miami Public Schools
Ms. Jill Douthit
601 16th Avenue, Northwest
Miami, OK 74354

Phone: (918) 533-4297

Phone: (918) 533-4297

Date Due: 01/15/26 17:00 (5 day TAT)

Received By: Jordan Anderson

Date Received: 01/08/26 16:13

Logged In By: Jordan Anderson

Date Logged In: 01/08/26 16:25

Samples Received at:	0.9°C				
Custody seals	No	Received on ice	Yes	Sufficient sample	Yes
Containers intact	Yes	Sample or temp blank frozen	No		
COC/Labels agree	Yes	Headspace in VOA vials	No		
Preservation confirmed	No	Correct containers	Yes		

Notes:

Preservation Confirmation

Container ID	Container Type	pH	Date/Time	Lot #
--------------	----------------	----	-----------	-------

Preservation Confirmed By

Date

Samples Labeled By

Date

Reviewed By

Date

(405) 488-2400
FAX: (405) 488-2404



PAGE: 1 OF 1
SAMPLE SERIES #: E6A10125

Page 7 of 7

[illegible]